







Tissue Processor Comparison Chart

						
	Intelsint ETP	Leica PELORIS	Sakura VIP 6	Sakura VIP 5	Leica ASP300 S	Dakewe HP300
Warranty	2-Years	1-Year	1-Year	1-Year	1-Year	2-Years
Full Service Contract	~\$8,000 annually	~ \$15,000 annually	~ \$12,000 annually	~ \$7,000 annually	~ \$7,000 annually	~ \$7,000 annually
Max cassettes per run	360	600	300	300	300	300
Processing Time (1mm Biopsies)	8 Hours (ETP) 45 Minutes (EFTP)	45 Minutes	8 Hours	8 Hours	8 Hours	8 Hours
Xylene Free Processing	Yes	Yes	Yes	Yes	Yes	Yes
Fume Control	Self-Contained (Carbon) + Retort Fume Extraction	Self-Contained (Carbon)	Self-Contained (Carbon)	Self-Contained (Carbon)	Self-Contained (Carbon)	Self-Contained (Carbon) + Hose Connection
Reagent Management	Automatic Optional Light Indication Optional	Automatic	Yes; User Defined	Yes; User Defined	Yes; User Defined	Automatic Light Indication
User Interface	Touchscreen	Touchscreen	Touchscreen	LCD Screen with buttons	Touchscreen	Touchscreen + Mobile App
Retorts	1	2	1	1	1	1
Bottle Size	5L	3.8L	3.7L	4.5L	4.3L	4.8L
Cassette Baskets	3	4	2	2	3	2
Reagent Stations	10	16	10	10	13	10
Paraffin Reservoirs	3 (5L)	4	4 (3 Small, 1 Large)	4 (3.7L)	3	4 (3 x 4.2L, 1 x 5.6L)
Dimensions (Inches)	27 W x 27.55 D x 59 H	33.7 W x 28.4 D x 59 H	24 W x 27 D x 54 H	24 W x 24 D x 54 H	24 W x 27 D x 52 H	24 W x 25 D x 46 H
Weight	498 lbs	550 lbs	352 lbs	252 lbs	308 lbs	396 lbs